

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL062422\
 Data File : VL039337.D
 Acq On : 24 Jun 2022 11:38
 Operator : SY/AP
 Sample : VL0624ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0624ABL01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/27/2022
 Supervised By :Mahesh Dadoda 06/30/2022

Quant Time: Jun 27 01:49:21 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL053122AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jun 01 01:22:18 2022
 QLast Update : Wed Jun 01 01:22:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.661	49	1163940	10.000	ppbv	-0.01
33) 1,4-Difluorobenzene	7.169	114	3363127	10.000	ppbv	-0.01
55) Chlorobenzene-d5	12.076	117	2946510	10.000	ppbv	-0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.404	95	2179116	8.946	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	89.500%
Target Compounds						
9) Propene	3.015	41	9009m	0.140	ppbv	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL062422\
Data File : VL039337.D
Acq On : 24 Jun 2022 11:38
Operator : SY/AP
Sample : VL0624ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0624ABL01

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 06/27/2022
Supervised By : Mahesh Dadoda 06/30/2022

Quant Time: Jun 27 01:49:21 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL053122AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jun
QLast Update : Wed Jun 01 01:22:18 2022
Response via : Initial Calibration

