

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL062422\
 Data File : VL039358.D
 Acq On : 25 Jun 2022 2:30
 Operator : SY/AP
 Sample : N3453-01DL 20X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SG-1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/27/2022
 Supervised By :Mahesh Dadoda 06/30/2022

Quant Time: Jun 27 01:55:02 2022

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL053122AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jun 01 01:22:18 2022

QLast Update : Wed Jun 01 01:22:18 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.661	49	1117566	10.000	ppbv	-0.01
33) 1,4-Difluorobenzene	7.166	114	3097953	10.000	ppbv	-0.02
55) Chlorobenzene-d5	12.076	117	2662791	10.000	ppbv	-0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.404	95	1952708	8.870	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	88.700%
Target Compounds						
						Qvalue
3) Chlorodifluoromethane	2.989	51	21807	0.153	ppbv	98
4) Chloromethane	3.144	50	11039m	0.151	ppbv	
9) Propene	3.022	41	17587m	0.284	ppbv	
13) Ethanol	3.594	45	211489	62.545	ppbv #	100
15) Acetone	3.845	43	608951	5.231	ppbv	99
19) Isopropyl Alcohol	3.964	45	177913	3.258	ppbv #	91
20) Methylene Chloride	4.343	84	42257m	0.861	ppbv	
26) Hexane	5.677	57	28378m	0.246	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL062422\
Data File : VL039358.D
Acq On : 25 Jun 2022 2:30
Operator : SY/AP
Sample : N3453-01DL 20X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SG-1

Quant Time: Jun 27 01:55:02 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL053122AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jun
QLast Update : Wed Jun 01 01:22:18 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/27/2022
Supervised By :Mahesh Dadoda 06/30/2022

