


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL062619\
 Data File : VL033963.D
 Acq On : 26 Jun 2019 9:48
 Operator : SY/AP
 Sample : BFB
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL060719AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Jun 07 08:53:27 2019
 Last Update : Fri Jun 07 08:53:27 2019



AutoFind: Scans 3608, 3609, 3610; Background Corrected with Scan 3585

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	23.8	184021	PASS
75	95	30	66	58.2	450432	PASS
95	95	100	100	100.0	773418	PASS
96	95	5	9	6.2	48114	PASS
173	174	0.00	2	0.8	4132	PASS
174	95	50	120	65.4	505621	PASS
175	174	4	9	7.3	36690	PASS
176	174	93	101	94.7	478848	PASS
177	176	5	9	6.8	32578	PASS