

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL062725\
 Data File : VL042676.D
 Acq On : 27 Jun 2025 12:57
 Operator : SY/MD
 Sample : Q2432-01DL 100X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV1DL

Quant Time: Jun 27 23:04:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL052925AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 30 02:01:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.780	49	110587	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.949	114	292442	10.000	ppbv	-0.01
55) Chlorobenzene-d5	8.872	117	252771	10.000	ppbv	-0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.370	95	195816	10.334	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.300%
Target Compounds						
					Qvalue	
13) Ethanol	1.719	45	1139m	1.268	ppbv	
15) Acetone	1.832	43	9339	0.701	ppbv	99
20) Methylene Chloride	2.059	84	13086	2.977	ppbv	96
26) Hexane	2.822	57	21578	1.303	ppbv #	41
34) 2-Butanone	2.557	43	4064m	0.195	ppbv	
52) Tetrachloroethene	8.215	164	46894	4.358	ppbv	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL062725\
Data File : VL042676.D
Acq On : 27 Jun 2025 12:57
Operator : SY/MD
Sample : Q2432-01DL 100X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV1DL

Quant Time: Jun 27 23:04:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL052925AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri May 30 02:01:56 2025
Response via : Initial Calibration

