

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL062921\
 Data File : VL037247.D
 Acq On : 29 Jun 2021 13:35
 Operator : SY/AP
 Sample : QC062221 CAN 10742
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC062221 CAN 10742

Quant Time: Jun 29 14:12:01 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL061621AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LWed Jun 16 14:06:32 2021
 QLast Update : Wed Jun 16 14:06:32 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1664612	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.16	114	4750945	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.07	117	4169968	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.40	95	3082522	9.76	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL062921\
Data File : VL037247.D
Acq On : 29 Jun 2021 13:35
Operator : SY/AP
Sample : OC062221 CAN 10742
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC062221 CAN 10742

Quant Time: Jun 29 14:12:01 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL061621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jun 16 14:06:32 2021
QLast Update : Wed Jun 16 14:06:32 2021
Response via : Initial Calibration

