

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL062921\
 Data File : VL037258.D
 Acq On : 29 Jun 2021 20:49
 Operator : SY/AP
 Sample : M2887-05DL 40X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV-5DL

Manual Integrations
 APPROVED

MMDadoda
 7/5/2021 3:38:49 PM

Quant Time: Jun 30 06:52:29 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL061621AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Wed Jun 1
 QLast Update : Wed Jun 16 14:06:32 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1615033	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.17	114	4770832	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.08	117	4232364	10.00	ppbv	-0.02

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.41	95	3145394	9.81	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.01	41	1060150	10.558	ppbv	94
10) Heptane	8.11	43	42942	0.177	ppbv #	77
13) Ethanol	3.62	45	16762m	1.079	ppbv	
15) Acetone	3.85	43	477659	2.623	ppbv	94
20) Methylene Chloride	4.34	84	287306	2.635	ppbv	93
26) Hexane	5.68	57	237941	1.194	ppbv #	43
34) 2-Butanone	5.26	43	187321m	1.010	ppbv	
36) Benzene	6.88	78	98395	0.255	ppbv	99
51) 2-Hexanone	10.15	43	16050m	0.129	ppbv	
53) Toluene	9.80	91	89457m	0.194	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL062921\
Data File : VL037258.D
Acq On : 29 Jun 2021 20:49
Operator : SY/AP
Sample : M2887-05DL 40X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV-5DL

Manual Integrations
APPROVED

MMDadoda
7/5/2021 3:38:49 PM

Quant Time: Jun 30 06:52:29 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL061621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jun
QLast Update : Wed Jun 16 14:06:32 2021
Response via : Initial Calibration

