

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL062921\
 Data File : VL037263.D
 Acq On : 30 Jun 2021 00:06
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

MMDadoda
 7/5/2021 3:39:10 PM

Quant Time: Jun 30 07:11:27 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL061621AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jun 1
 QLast Update : Wed Jun 16 14:06:32 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1597217	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.17	114	4688056	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.08	117	4089188	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.41	95	3046968	9.83	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.87	43	55034m	0.306	ppbv	
20) Methylene Chloride	4.34	84	28457m	0.264	ppbv	
34) 2-Butanone	5.27	43	19388m	0.106	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL062921\
Data File : VL037263.D
Acq On : 30 Jun 2021 00:06
Operator : SY/AP
Sample : VIBLK
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

MMDadoda
7/5/2021 3:39:10 PM

Quant Time: Jun 30 07:11:27 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL061621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jun
QLast Update : Wed Jun 16 14:06:32 2021
Response via : Initial Calibration

