

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL070221\
 Data File : VL037276.D
 Acq On : 2 Jul 2021 15:49
 Operator : SY/AP
 Sample : QC063021 CAN 10197
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC063021 CAN 10197

Quant Time: Jul 04 22:59:07 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL061621AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jun 16 14:06:32 2021
 QLast Update : Wed Jun 16 14:06:32 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	1241641	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	3602152	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.11	117	3110428	10.00	ppbv	0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.44	95	2359747	10.01	ppbv	0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.10%


Target Compounds	Ovalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL070221\
Data File : VL037276.D
Acq On : 2 Jul 2021 15:49
Operator : SY/AP
Sample : OC063021 CAN 10197
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC063021 CAN 10197

Quant Time: Jul 04 22:59:07 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL061621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jun 16 14:06:32 2021
QLast Update : Wed Jun 16 14:06:32 2021
Response via : Initial Calibration





#1

Bromochloromethane

Concen: 10.000 ppbv

RT: 5.68 min Scan# 901

Delta R.T. 0.01 min

Lab File: VL037276.D

Acq: 2 Jul 2021 15:49

Instrument :

MSVOA_L

ClientSampleId :

QC063021 CAN 10197


Tot Ion: 49 Resp: 1241641

Ion Ratio Lower Upper

49 100

128 55.1 29.3 88.0




130 72.2 35.9 107.8



Abundance Ion 49.10 (48.80 to 49.80): VL0

Ion 128.00 (127.70 to 128.70): VL0

Ion 130.00 (129.70 to 130.70): VL0



#33

1,4-Difluorobenzene

Concen: 10.000 ppbv

RT: 7.19 min Scan# 1370

Delta R.T. 0.01 min

Lab File: VL037276.D

Acq: 2 Jul 2021 15:49


Tgt Ion: 114 Resp: 3602152

Ion Ratio Lower Upper

114 100

63 25.1 19.6 29.4



88 17.8 14.2 21.2




Abundance Ion 114.10 (113.80 to 114.80): VL0

Ion 63.10 (62.80 to 63.80): VL0

Ion 88.10 (87.80 to 88.80): VL0









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 12.11 min Scan# 2899
Delta R.T. 0.01 min
Lab File: VL037276.D
Acq: 2 Jul 2021 15:49

Instrument :
MSVOA_L
ClientSampleId :
QC063021 CAN 10197

Tot Ion:117 Resp: 3110428
Ion Ratio Lower Upper
117 100
82 67.2 51.7 77.5
119 35.5 47.6 71.4#



#68
1-Bromo-4-Fluorobenzene
Concen: 10.012 ppbv
RT: 14.44 min Scan# 3624
Delta R.T. 0.01 min
Lab File: VL037276.D
Acq: 2 Jul 2021 15:49

Tgt Ion: 95 Resp: 2359747
Ion Ratio Lower Upper
95 100
174 74.6 57.0 85.4
175 5.3 4.4 6.6

