

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL070522\
Data File : VL039420.D
Acq On : 5 Jul 2022 15:04
Operator : SY/AP
Sample : VL0705ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0705ABL01

Quant Time: Jul 06 02:36:30 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL062922AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jul 06 02:15:38 2022
QLast Update : Wed Jul 06 02:15:38 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.671	49	1192052	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	7.179	114	3395213	10.000	ppbv	0.02
55) Chlorobenzene-d5	12.089	117	2865007	10.000	ppbv	0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.417	95	2135400	10.047	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL070522\
Data File : VL039420.D
Acq On : 5 Jul 2022 15:04
Operator : SY/AP
Sample : VL0705ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0705ABL01

Quant Time: Jul 06 02:36:30 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL062922AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jul 06 02:15:38 2022
QLast Update : Wed Jul 06 02:15:38 2022
Response via : Initial Calibration

