

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL070620\
 Data File : VL035153.D
 Acq On : 6 Jul 2020 19:47
 Operator : SY/AP
 Sample : L3190-03 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 EXHAUST-02

Manual Integrations
 APPROVED

MMDadoda
 7/9/2020 12:00:55 PM

Quant Time: Jul 07 05:52:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL061920AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Sat Jun 20 2020
 QLast Update : Sat Jun 20 02:10:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1013123	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	2385918	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.10	117	2227338	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.44	95	1882122	10.17	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.70%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	2.97	41	23821	0.538	ppbv	91
13) Ethanol	3.57	45	22160	1.324	ppbv	99
15) Acetone	3.82	43	293458	2.396	ppbv	92
17) tert-Butyl alcohol	4.29	59	34041m	0.314	ppbv	
19) Isopropyl Alcohol	3.95	45	13657m	0.172	ppbv	
20) Methylene Chloride	4.31	84	193519	3.904	ppbv	95
26) Hexane	5.66	57	62122	0.623	ppbv #	44
27) Carbon Disulfide	4.52	76	99122	0.961	ppbv	95
36) Benzene	6.88	78	67540	0.366	ppbv #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL070620\
Data File : VL035153.D
Acq On : 6 Jul 2020 19:47
Operator : SY/AP
Sample : L3190-03 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
EXHAUST-02

Manual Integrations
APPROVED

MMDadoda
7/9/2020 12:00:55 PM

Quant Time: Jul 07 05:52:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL061920AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Sat Jun
QLast Update : Sat Jun 20 02:10:36 2020
Response via : Initial Calibration

