

Data Path : W:\HPCHEM1\MSVOA L\Data\VL070715\
 Data File : VL025609.D
 Acq On : 7 Jul 2015 13:52
 Operator : SY/MD
 Sample : VSTDIC0.03
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.03

Manual Integrations
 APPROVED

sam
 7/8/2015 11:06:02 AM

Quant Time: Jul 08 01:19:40 2015
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL070715AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Jul 0
 QLast Update : Tue Jul 07 13:08:42 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	6.65	49	550814	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	8.32	114	1448251	10.00	ppbv	0.00
55) Chlorobenzene-d5	13.74	117	1876970	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	16.29	95	1621576	10.46	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.89	62	1050	0.03	ppbv #	43
32) 1,1,1-Trichloroethane	7.59	97	4425m	0.03	ppbv	
35) Carbon Tetrachloride	8.16	117	4791	0.04	ppbv	87
38) Trichloroethene	9.06	130	1903	0.03	ppbv #	80
52) Tetrachloroethene	12.79	164	1739	0.03	ppbv	94
63) 1,1,2,2-Tetrachloroethane	15.49	83	6364	0.03	ppbv #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\Data\VL070715\
Data File : VL025609.D
Acq On : 7 Jul 2015 13:52
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Manual Integrations
APPROVED

sam
7/8/2015 11:06:02 AM

Quant Time: Jul 08 01:19:40 2015
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL070715AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Jul
QLast Update : Tue Jul 07 13:08:42 2015
Response via : Initial Calibration

