

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL070715\
 Data File : VL025614.D
 Acq On : 7 Jul 2015 19:18
 Operator : SY/MD
 Sample : G2653-01
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 AIR-LOD

Manual Integrations
 APPROVED

sam
 7/8/2015 11:06:14 AM

Quant Time: Jul 08 11:26:55 2015
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL070715AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LWed Jul 08 12:54:17 2015
 QLast Update : Wed Jul 08 02:22:55 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	6.65	49	567018	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	8.33	114	1541589	10.00	ppbv	0.00
55) Chlorobenzene-d5	13.75	117	1972132	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	16.29	95	1776052m	10.68	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	106.80%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	4.37	45	1559m	0.38	ppbv	
40) 1,4-Dioxane	9.21	88	1204	0.14	ppbv #	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL070715\
Data File : VL025614.D
Acq On : 7 Jul 2015 19:18
Operator : SY/MD
Sample : G2653-01
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
AIR-LOD

Manual Integrations
APPROVED

sam
7/8/2015 11:06:14 AM

Quant Time: Jul 08 11:26:55 2015
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL070715AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jul
QLast Update : Wed Jul 08 02:22:55 2015
Response via : Initial Calibration

