

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL070819\
 Data File : VL034008.D
 Acq On : 8 Jul 2019 11:05
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Jul 09 04:03:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL070219AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Jul 02 19:31:31 2019
 QLast Update : Tue Jul 02 19:31:31 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	2.064	1.485	28.1	82	0.00
3	Chlorodifluoromethane	1.235	1.092	11.6	92	0.00
4	Chloromethane	0.710	0.712	-0.3	106	0.00
5 T	Vinyl Chloride	0.798	0.775	2.9	105	0.00
6 T	Bromomethane	0.471	0.486	-3.2	103	0.00
7	Chloroethane	0.285	0.292	-2.5	103	0.00
8 T	Dichlorotetrafluoroethane	2.012	1.964	2.4	103	0.00
9 T	Propene	0.454	0.421	7.3	95	0.00
10 T	Heptane	1.137	1.191	-4.7	93	0.00
11 T	Trichlorofluoromethane	2.321	2.105	9.3	101	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.642	1.526	7.1	103	0.00
13	Ethanol	0.141	0.121	14.2	99	0.00
14 T	Bromoethene	0.613	0.653	-6.5	104	0.00
15 T	Acetone	1.681	1.418	15.6	101	0.00
16 T	1,3-Butadiene	0.594	0.632	-6.4	103	0.00
17	tert-Butyl alcohol	1.340	1.270	5.2	96	0.00
18 T	1,1-Dichloroethene	0.732	0.721	1.5	104	0.00
19 T	Isopropyl Alcohol	1.129	0.990	12.3	95	0.00
20 T	Methylene Chloride	0.733	0.612	16.5	102	0.00
21 T	Allyl Chloride	1.048	1.055	-0.7	105	0.00
22 T	trans-1,2-Dichloroethene	0.697	0.722	-3.6	102	0.00
23 T	Vinyl Acetate	1.678	1.833	-9.2	99	0.00
24 T	1,1-Dichloroethane	1.487	1.463	1.6	99	0.00
25 T	Ethyl Acetate	2.300	2.366	-2.9	99	0.00
26 T	Hexane	1.052	1.085	-3.1	101	0.00
27 T	Carbon Disulfide	1.789	1.864	-4.2	104	0.00
28 T	Methyl tert-Butyl Ether	1.782	1.899	-6.6	98	0.00
29 T	Chloroform	1.978	1.862	5.9	98	0.00
30 T	Cyclohexane	0.813	0.914	-12.4	101	0.00
31 T	cis-1,2-Dichloroethene	1.017	1.110	-9.1	98	0.00
32 T	1,1,1-Trichloroethane	2.004	1.856	7.4	96	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
34 T	2-Butanone	0.529	0.562	-6.2	98	0.00
35 T	Carbon Tetrachloride	0.792	0.795	-0.4	99	0.00
36 T	Benzene	0.782	0.859	-9.8	97	0.00
37 T	1,2-Dichloroethane	0.576	0.565	1.9	94	0.00
38 T	Trichloroethene	0.348	0.375	-7.8	97	0.00
39 T	1,2-Dichloropropane	0.291	0.307	-5.5	96	0.00
40 T	1,4-Dioxane	0.134	0.136	-1.5	98	0.00
41 T	Tetrahydrofuran	0.260	0.296	-13.8	97	0.00
42 T	Bromodichloromethane	0.773	0.808	-4.5	98	0.00
43	Methyl Methacrylate	0.297	0.346	-16.5	98	0.00
44 T	2,2,4-Trimethylpentane	1.351	1.423	-5.3	96	0.00
45 T	t-1,3-Dichloropropene	0.388	0.490	-26.3	97	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL070819\
 Data File : VL034008.D
 Acq On : 8 Jul 2019 11:05
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Jul 09 04:03:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL070219AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Jul 02 19:31:31 2019
 QLast Update : Tue Jul 02 19:31:31 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	cis-1,3-Dichloropropene	0.450	0.536	-19.1	96	0.00
47 T	1,1,2-Trichloroethane	0.366	0.389	-6.3	101	0.00
48 T	Dibromochloromethane	0.698	0.709	-1.6	98	0.00
49 T	Bromoform	0.596	0.671	-12.6	102	0.00
50 T	4-Methyl-2-Pentanone	0.785	0.849	-8.2	96	0.00
51 T	2-Hexanone	0.633	0.711	-12.3	98	0.00
52 T	Tetrachloroethene	0.337	0.368	-9.2	98	0.00
53 T	Toluene	0.896	1.077	-20.2	99	0.00
54 T	1,2-Dibromoethane	0.540	0.605	-12.0	101	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	103	0.00
56	1,1,1,2-Tetrachloroethane	0.447	0.429	4.0	102	0.00
57 T	Chlorobenzene	0.796	0.752	5.5	102	0.00
58 T	Ethyl Benzene	1.203	1.312	-9.1	101	0.00
59 T	m/p-Xylene	1.089	1.087	0.2	100	0.00
60 T	o-Xylene	1.089	1.156	-6.2	107	0.00
61 T	Styrene	0.721	0.875	-21.4	108	0.00
62	Isopropylbenzene	1.675	1.553	7.3	80	0.00
63 T	1,1,2,2-Tetrachloroethane	0.843	0.803	4.7	106	0.00
64	n-propylbenzene	0.434	0.395	9.0	79	0.00
65	tert-Butylbenzene	1.424	1.323	7.1	75	0.00
66 T	Benzyl Chloride	0.729	0.743	-1.9	75	0.00
67	sec-Butylbenzene	1.999	1.851	7.4	77	0.00
68 S	1-Bromo-4-Fluorobenzene	0.853	0.782	8.3	81	0.00
69	p-Isopropyltoluene	1.648	1.539	6.6	77	0.00
70	n-Butylbenzene	1.731	1.591	8.1	76	0.00
71	2-Chlorotoluene	1.225	1.171	4.4	79	0.00
72 T	4-Ethyltoluene	1.399	1.275	8.9	77	0.00
73 T	1,3,5-Trimethylbenzene	1.302	1.217	6.5	75	0.00
74 T	1,2,4-Trimethylbenzene	1.442	1.287	10.7	75	0.00
75 T	1,3-Dichlorobenzene	0.916	0.763	16.7	75	0.00
76 T	1,4-Dichlorobenzene	0.862	0.763	11.5	77	0.00
77 T	1,2-Dichlorobenzene	0.848	0.735	13.3	76	0.00
78 T	Hexachloro-1,3-Butadiene	0.722	0.719	0.4	100	0.00
79 T	Naphthalene	1.212	1.205	0.6	79	0.00
80 T	Naphthalene,2-methyl-	0.441	0.680	-54.2#	130	0.00
81 T	1,2,4-Trichlorobenzene	0.778	0.717	7.8	78	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0