

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL071318\
Data File : VL032255.D
Acq On : 13 Jul 2018 21:27
Operator : SY/AP
Sample : VIBLK
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Jul 14 04:01:56 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL071218AIR.M
Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Jul 13 04:25:14 2018
QLast Update : Fri Jul 13 04:25:14 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.84	49	1509113	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.37	114	3403145	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.33	117	3784535	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.67	95	2984359	10.19	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.90%

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL071318\
Data File : VL032255.D
Acq On : 13 Jul 2018 21:27
Operator : SY/AP
Sample : VIBLK
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Jul 14 04:01:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL071218AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jul 13 04:25:14 2018
QLast Update : Fri Jul 13 04:25:14 2018
Response via : Initial Calibration

