

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL071322\
 Data File : VL039473.D
 Acq On : 13 Jul 2022 20:35
 Operator : SY/AP
 Sample : QC071122 CAN 10599
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC071122 CAN 10599

Quant Time: Jul 14 01:25:03 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL062922AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jul 6
 QLast Update : Wed Jul 06 02:15:38 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 07/14/2022
 Supervised By :Mahesh Dadoda 07/14/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.677	49	1157553	10.000	ppbv	0.03
33) 1,4-Difluorobenzene	7.182	114	3199765	10.000	ppbv	0.03
55) Chlorobenzene-d5	12.092	117	2844044	10.000	ppbv	0.03
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.423	95	2185118	10.357	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.600%
Target Compounds						
9) Propene	3.031	41	9386m	0.142	ppbv	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL071322\
Data File : VL039473.D
Acq On : 13 Jul 2022 20:35
Operator : SY/AP
Sample : QC071122 CAN 10599
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC071122 CAN 10599

Quant Time: Jul 14 01:25:03 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL062922AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jul
QLast Update : Wed Jul 06 02:15:38 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 07/14/2022
Supervised By :Mahesh Dadoda 07/14/2022

