

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL071719\
 Data File : VL034041.D
 Acq On : 17 Jul 2019 21:12
 Operator : SY/AP
 Sample : K3804-01DL2 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SG-1DL2

Manual Integrations
 APPROVED

MMDadoda
 7/19/2019 1:37:21 AM

Quant Time: Jul 18 05:37:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL070219AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Jul 0
 QLast Update : Tue Jul 02 19:31:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.72	49	1003866	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.25	114	3977733	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.19	117	2912367	10.00	ppbv	-0.02

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.52	95	2276958	9.17	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	91.70%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.02	41	55284	1.212	ppbv	95
13) Ethanol	3.63	45	620958	43.832	ppbv	93
15) Acetone	3.88	43	345096	2.045	ppbv	96
20) Methylene Chloride	4.37	84	8571m	0.116	ppbv	
27) Carbon Disulfide	4.59	76	236146	1.315	ppbv	95
34) 2-Butanone	5.30	43	23349m	0.111	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL071719\
Data File : VL034041.D
Acq On : 17 Jul 2019 21:12
Operator : SY/AP
Sample : K3804-01DL2 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SG-1DL2

Manual Integrations
APPROVED

MMDadoda
7/19/2019 1:37:21 AM

Quant Time: Jul 18 05:37:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL070219AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Jul
QLast Update : Tue Jul 02 19:31:31 2019
Response via : Initial Calibration

