

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL071818\
 Data File : VL032263.D
 Acq On : 18 Jul 2018 11:55
 Operator : SY/AP
 Sample : VSTDICC0.1
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.1

Manual Integrations
 APPROVED

MMDadoda
 7/20/2018 2:53:07 PM

Quant Time: Jul 18 12:52:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL071818AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jul 18 12:52:55 2018
 QLast Update : Wed Jul 18 09:58:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.84	49	1724157	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.37	114	3620249	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.33	117	3978432	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.67	95	3354558	10.89	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	108.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.34	62	10705	0.09	ppbv	93
32) 1,1,1-Trichloroethane	6.69	97	24419m	0.09	ppbv	
35) Carbon Tetrachloride	7.22	117	23664	0.10	ppbv #	88
38) Trichloroethene	8.05	130	11429m	0.09	ppbv	
52) Tetrachloroethene	11.46	164	12890m	0.10	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.93	83	41787m	0.11	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL071818\
Data File : VL032263.D
Acq On : 18 Jul 2018 11:55
Operator : SY/AP
Sample : VSTDICC0.1
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Manual Integrations
APPROVED

MMDadoda
7/20/2018 2:53:07 PM

Quant Time: Jul 18 12:52:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL071818AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jul
QLast Update : Wed Jul 18 09:58:24 2018
Response via : Initial Calibration

