

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL071818\
 Data File : VL032267.D
 Acq On : 18 Jul 2018 19:08
 Operator : SY/AP
 Sample : VL0718ABL01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0718ABL01

Manual Integrations
 APPROVED

MMDadoda
 7/20/2018 2:53:11 PM

Quant Time: Jul 20 14:00:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL071818AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jul 18 14:38:53 2018
 QLast Update : Wed Jul 18 14:38:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.84	49	1711963	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.37	114	3603285	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.33	117	4220978	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.67	95	3344247	9.77	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.70%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	4.00	43	25833	0.14	ppbv	# 82
20) Methylene Chloride	4.48	84	10920m	0.16	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL071818\
Data File : VL032267.D
Acq On : 18 Jul 2018 19:08
Operator : SY/AP
Sample : VL0718ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0718ABL01

Manual Integrations
APPROVED

MMDadoda
7/20/2018 2:53:11 PM

Quant Time: Jul 20 14:00:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL071818AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jul
QLast Update : Wed Jul 18 14:38:53 2018
Response via : Initial Calibration

