

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL071922\
 Data File : VL039486.D
 Acq On : 19 Jul 2022 17:29
 Operator : SY/AP
 Sample : N3750-01DL 20X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV-1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 07/20/2022
 Supervised By :Mahesh Dadoda 07/21/2022

Quant Time: Jul 20 05:40:51 2022

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL062922AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jul 6 2022

QLast Update : Wed Jul 06 02:15:38 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.677	49	1252290	10.000	ppbv	0.03
33) 1,4-Difluorobenzene	7.189	114	3543189	10.000	ppbv	0.03
55) Chlorobenzene-d5	12.102	117	3173959	10.000	ppbv	0.04
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.433	95	2391562	10.157	ppbv	0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.600%
Target Compounds						
					Qvalue	
9) Propene	3.021	41	117096	1.632	ppbv	92
13) Ethanol	3.629	45	6240m	1.499	ppbv	
15) Acetone	3.864	43	803720	5.425	ppbv #	89
20) Methylene Chloride	4.359	84	83864	1.420	ppbv	87
26) Hexane	5.700	57	181056	1.479	ppbv #	38
29) Chloroform	5.761	83	425914	2.000	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL071922\
Data File : VL039486.D
Acq On : 19 Jul 2022 17:29
Operator : SY/AP
Sample : N3750-01DL 20X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV-1DL

Quant Time: Jul 20 05:40:51 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL062922AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jul
QLast Update : Wed Jul 06 02:15:38 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 07/20/2022
Supervised By : Mahesh Dadoda 07/21/2022

