

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL071922\
 Data File : VL039490.D
 Acq On : 19 Jul 2022 20:30
 Operator : SY/AP
 Sample : N3750-03DL 20X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV-3DL

Quant Time: Jul 20 05:41:36 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL062922AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jul 06 02:15:38 2022
 QLast Update : Wed Jul 06 02:15:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.677	49	1222217	10.000	ppbv	0.03
33) 1,4-Difluorobenzene	7.188	114	3442493	10.000	ppbv	0.03
55) Chlorobenzene-d5	12.101	117	3067019	10.000	ppbv	0.04
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.429	95	2291799	10.073	ppbv	0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	= 100.700%	
Target Compounds						
						Qvalue
9) Propene	3.021	41	64899	0.927	ppbv	91
13) Ethanol	3.626	45	3519	0.866	ppbv	68
15) Acetone	3.864	43	373375	2.582	ppbv #	89
20) Methylene Chloride	4.359	84	34545	0.599	ppbv	95
26) Hexane	5.700	57	284545	2.381	ppbv #	37

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL071922\
Data File : VL039490.D
Acq On : 19 Jul 2022 20:30
Operator : SY/AP
Sample : N3750-03DL 20X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV-3DL

Quant Time: Jul 20 05:41:36 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL062922AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jul 06 02:15:38 2022
QLast Update : Wed Jul 06 02:15:38 2022
Response via : Initial Calibration

