

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL071922\
 Data File : VL039495.D
 Acq On : 20 Jul 2022 00:13
 Operator : SY/AP
 Sample : N3750-05DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA-1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 07/20/2022
 Supervised By :Mahesh Dadoda 07/21/2022

Quant Time: Jul 20 05:42:39 2022

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL062922AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jul 6 2022

QLast Update : Wed Jul 06 02:15:38 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.674	49	1129322	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	7.185	114	3274022	10.000	ppbv	0.03
55) Chlorobenzene-d5	12.098	117	2940120	10.000	ppbv	0.03
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.426	95	2127672	9.755	ppbv	0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.500%
Target Compounds						
13) Ethanol	3.616	45	32617m	8.689	ppbv	Qvalue
15) Acetone	3.867	43	128148	0.959	ppbv #	89
20) Methylene Chloride	4.349	84	77434	1.454	ppbv	98
26) Hexane	5.700	57	67142	0.608	ppbv #	43

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL071922\
Data File : VL039495.D
Acq On : 20 Jul 2022 00:13
Operator : SY/AP
Sample : N3750-05DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA-1DL

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 07/20/2022
Supervised By : Mahesh Dadoda 07/21/2022

Quant Time: Jul 20 05:42:39 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL062922AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jul
QLast Update : Wed Jul 06 02:15:38 2022
Response via : Initial Calibration

