

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL071922\
 Data File : VL039504.D
 Acq On : 20 Jul 2022 8:09
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Quant Time: Jul 21 05:29:35 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL062922AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jul 06 02:15:38 2022
 QLast Update : Wed Jul 06 02:15:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.677	49	917505	10.000	ppbv	0.03
33) 1,4-Difluorobenzene	7.182	114	2594047	10.000	ppbv	0.03
55) Chlorobenzene-d5	12.095	117	2309890	10.000	ppbv	0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.423	95	1766579	10.309	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.100%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL071922\
Data File : VL039504.D
Acq On : 20 Jul 2022 8:09
Operator : SY/AP
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Jul 21 05:29:35 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL062922AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jul 06 02:15:38 2022
QLast Update : Wed Jul 06 02:15:38 2022
Response via : Initial Calibration









#1
 Bromochloromethane
 Concen: 10.000 ppbv
 RT: 5.677 min Scan# 891
 Delta R.T. 0.026 min
 Lab File: VL039504.D
 Acq: 20 Jul 2022 8:09




Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK


Tgt Ion: 49 Resp: 917505
 Ion Ratio Lower Upper
 49 100
 128 53.9 26.7 80.0
 130 71.8 34.3 102.8



#33
 1,4-Difluorobenzene
 Concen: 10.000 ppbv
 RT: 7.182 min Scan# 1367
 Delta R.T. 0.026 min
 Lab File: VL039504.D
 Acq: 20 Jul 2022 8:09

Tgt Ion: 114 Resp: 2594047
 Ion Ratio Lower Upper
 114 100
 63 25.4 20.2 30.4
 88 18.2 14.5 21.7





#55

Chlorobenzene-d5

Concen: 10.000 ppbv

RT: 12.095 min Scan# 2886

Delta R.T. 0.029 min

Lab File: VL039504.D

Acq: 20 Jul 2022 8:09

Instrument :

MSVOA_L

ClientSampleId :

VIBLK





Tgt Ion:117 Resp: 2309890

Ion Ratio Lower Upper

117 100

82 65.1 50.8 76.2

119 33.7 24.1 36.1



#68

1-Bromo-4-Fluorobenzene

Concen: 10.309 ppbv

RT: 14.423 min Scan# 3619

Delta R.T. 0.032 min

Lab File: VL039504.D

Acq: 20 Jul 2022 8:09

Tgt Ion: 95 Resp: 1766579

Ion Ratio Lower Upper

95 100

174 70.8 55.9 83.9

175 5.3 4.2 6.4

