

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL072120\
 Data File : VL035216.D
 Acq On : 21 Jul 2020 17:58
 Operator : SY/AP
 Sample : L3290-01DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 1ST-FLOOR-EAST-CLASSROOMDL

Manual Integrations
 APPROVED

MMDadoda
 7/23/2020 2:59:33 PM

Quant Time: Jul 22 01:36:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL070820AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jul 08 07:05:23 2020
 QLast Update : Wed Jul 08 07:05:23 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1057520	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	2580091	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.11	117	2412156	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.45	95	1966024	10.03	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	2.98	41	22078m	0.500	ppbv	
13) Ethanol	3.57	45	170385	12.123	ppbv	95
15) Acetone	3.83	43	155519	1.086	ppbv	97
19) Isopropyl Alcohol	3.94	45	67161	0.778	ppbv #	90
20) Methylene Chloride	4.32	84	82554	1.325	ppbv	90
26) Hexane	5.67	57	18644	0.190	ppbv #	42
53) Toluene	9.82	91	15993m	0.081	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL072120\
Data File : VL035216.D
Acq On : 21 Jul 2020 17:58
Operator : SY/AP
Sample : L3290-01DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
1ST-FLOOR-EAST-CLASSROOMDL

Manual Integrations
APPROVED

MMDadoda
7/23/2020 2:59:33 PM

Quant Time: Jul 22 01:36:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL070820AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jul
QLast Update : Wed Jul 08 07:05:23 2020
Response via : Initial Calibration

