

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072623\
 Data File : VL040556.D
 Acq On : 27 Jul 2023 1:25
 Operator : SY/MD
 Sample : VIBLK
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Quant Time: Jul 27 04:46:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL071023AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Thu Jul 27 04:45:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.176	49	1024372	10.000	ppbv	-0.03
33) 1,4-Difluorobenzene	6.639	114	2713903	10.000	ppbv	-0.03
55) Chlorobenzene-d5	11.455	117	2314504	10.000	ppbv	-0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.767	95	1823986	9.957	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.600%
Target Compounds						
					Qvalue	
15) Acetone	3.478	43	15652	0.133	ppbv #	58
20) Methylene Chloride	3.915	84	8248	0.164	ppbv #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072623\
Data File : VL040556.D
Acq On : 27 Jul 2023 1:25
Operator : SY/MD
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Jul 27 04:46:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL071023AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Thu Jul 27 04:45:44 2023
Response via : Initial Calibration

