

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL073019\
Data File : VL034056.D
Acq On : 30 Jul 2019 8:41
Operator : SY/AP
Sample : VL0730ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0730ABL01

Quant Time: Jul 31 02:14:58 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL073019AIR.M
Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LTue Jul 30 06:55:18 2019
QLast Update : Tue Jul 30 06:55:18 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.71	49	1014894	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.24	114	4019837	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.18	117	2843407	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.52	95	2355874	10.48	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.80%

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL073019\
Data File : VL034056.D
Acq On : 30 Jul 2019 8:41
Operator : SY/AP
Sample : VL0730ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0730ABL01

Quant Time: Jul 31 02:14:58 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL073019AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Jul 30 06:55:18 2019
QLast Update : Tue Jul 30 06:55:18 2019
Response via : Initial Calibration

