

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL073020\
 Data File : VL035264.D
 Acq On : 30 Jul 2020 2:47
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Jul 30 07:50:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL072920AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jul 29 07:14:29 2020
 QLast Update : Wed Jul 29 07:14:29 2020
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	1.194	1.067	10.6	92	0.00
3	Chlorodifluoromethane	1.107	0.980	11.5	93	0.00
4	Chloromethane	0.623	0.570	8.5	95	0.00
5 T	Vinyl Chloride	0.649	0.576	11.2	92	0.00
6 T	Bromomethane	0.401	0.363	9.5	92	0.00
7	Chloroethane	0.234	0.221	5.6	92	0.00
8 T	Dichlorotetrafluoroethane	1.515	1.361	10.2	93	0.00
9 T	Propene	0.386	0.377	2.3	92	0.00
10 T	Heptane	1.044	1.120	-7.3	92	0.00
11 T	Trichlorofluoromethane	1.606	1.453	9.5	92	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.139	1.008	11.5	91	0.00
13	Ethanol	0.157	0.092	41.4#	84	0.00
14 T	Bromoethene	0.465	0.440	5.4	92	0.00
15 T	Acetone	1.334	1.090	18.3	93	0.00
16 T	1,3-Butadiene	0.520	0.514	1.2	94	0.00
17	tert-Butyl alcohol	1.287	1.203	6.5	94	0.00
18 T	1,1-Dichloroethene	0.506	0.460	9.1	89	0.00
19 T	Isopropyl Alcohol	0.880	0.799	9.2	96	0.00
20 T	Methylene Chloride	0.607	0.440	27.5	95	0.00
21 T	Allyl Chloride	0.760	0.740	2.6	91	0.00
22 T	trans-1,2-Dichloroethene	0.486	0.453	6.8	91	0.00
23 T	Vinyl Acetate	1.340	1.372	-2.4	93	0.00
24 T	1,1-Dichloroethane	1.250	1.171	6.3	93	0.00
25 T	Ethyl Acetate	2.083	1.980	4.9	92	0.00
26 T	Hexane	0.940	0.907	3.5	94	0.00
27 T	Carbon Disulfide	1.270	1.196	5.8	89	0.00
28 T	Methyl tert-Butyl Ether	1.343	1.370	-2.0	93	0.00
29 T	Chloroform	1.576	1.452	7.9	93	0.00
30 T	Cyclohexane	0.672	0.749	-11.5	92	0.00
31 T	cis-1,2-Dichloroethene	0.819	0.872	-6.5	93	0.00
32 T	1,1,1-Trichloroethane	1.613	1.434	11.1	92	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	90	0.00
34 T	2-Butanone	0.509	0.499	2.0	91	0.00
35 T	Carbon Tetrachloride	0.670	0.625	6.7	91	0.00
36 T	Benzene	0.695	0.729	-4.9	93	0.00
37 T	1,2-Dichloroethane	0.486	0.470	3.3	92	0.00
38 T	Trichloroethene	0.288	0.306	-6.3	92	0.00
39 T	1,2-Dichloropropane	0.277	0.288	-4.0	92	0.00
40 T	1,4-Dioxane	0.109	0.120	-10.1	92	0.00
41 T	Tetrahydrofuran	0.249	0.272	-9.2	92	0.00
42 T	Bromodichloromethane	0.648	0.652	-0.6	91	0.00
43	Methyl Methacrylate	0.301	0.336	-11.6	90	0.00
44 T	2,2,4-Trimethylpentane	1.225	1.284	-4.8	93	0.00
45 T	t-1,3-Dichloropropene	0.303	0.367	-21.1	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL073020\
 Data File : VL035264.D
 Acq On : 30 Jul 2020 2:47
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Jul 30 07:50:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL072920AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jul 29 07:14:29 2020
 QLast Update : Wed Jul 29 07:14:29 2020
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	cis-1,3-Dichloropropene	0.370	0.437	-18.1	90	0.00
47 T	1,1,2-Trichloroethane	0.328	0.321	2.1	91	0.00
48 T	Dibromochloromethane	0.516	0.545	-5.6	91	0.00
49 T	Bromoform	0.464	0.493	-6.2	89	0.00
50 T	4-Methyl-2-Pentanone	0.737	0.791	-7.3	91	0.00
51 T	2-Hexanone	0.559	0.642	-14.8	90	0.00
52 T	Tetrachloroethene	0.265	0.274	-3.4	91	0.00
53 T	Toluene	0.739	0.870	-17.7	92	0.00
54 T	1,2-Dibromoethane	0.474	0.486	-2.5	91	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
56	1,1,1,2-Tetrachloroethane	0.460	0.423	8.0	90	0.00
57 T	Chlorobenzene	0.752	0.702	6.6	91	0.00
58 T	Ethyl Benzene	1.059	1.204	-13.7	92	0.00
59 T	m/p-Xylene	0.986	1.020	-3.4	91	0.00
60 T	o-Xylene	0.984	1.025	-4.2	91	0.00
61 T	Styrene	0.597	0.717	-20.1	91	0.00
62	Isopropylbenzene	1.464	1.508	-3.0	92	0.00
63 T	1,1,2,2-Tetrachloroethane	0.846	0.747	11.7	90	0.00
64	n-propylbenzene	0.389	0.414	-6.4	94	0.00
65	tert-Butylbenzene	1.326	1.348	-1.7	92	0.00
66 T	Benzyl Chloride	0.520	0.520	0.0	84	0.00
67	sec-Butylbenzene	1.938	1.924	0.7	92	0.00
68 S	1-Bromo-4-Fluorobenzene	0.799	0.773	3.3	89	0.00
69	p-Isopropyltoluene	1.581	1.610	-1.8	90	0.00
70	n-Butylbenzene	1.710	1.701	0.5	91	0.00
71	2-Chlorotoluene	1.104	1.169	-5.9	91	0.00
72 T	4-Ethyltoluene	1.107	1.225	-10.7	91	0.00
73 T	1,3,5-Trimethylbenzene	1.080	1.150	-6.5	92	0.00
74 T	1,2,4-Trimethylbenzene	1.257	1.210	3.7	90	0.00
75 T	1,3-Dichlorobenzene	0.812	0.713	12.2	90	0.00
76 T	1,4-Dichlorobenzene	0.734	0.703	4.2	89	0.00
77 T	1,2-Dichlorobenzene	0.727	0.680	6.5	89	0.00
78 T	Hexachloro-1,3-Butadiene	0.595	0.453	23.9	84	0.00
79 T	Naphthalene	1.000	0.954	4.6	79	0.00
80 T	Naphthalene,2-methyl-	0.323	0.290	10.2	67	0.00
81 T	1,2,4-Trichlorobenzene	0.608	0.554	8.9	82	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0