

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL073020\
 Data File : VL035270.D
 Acq On : 30 Jul 2020 7:29
 Operator : SY/AP
 Sample : OC CAN 10601
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC CAN 10601

Quant Time: Aug 02 15:37:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL072920AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jul 29 07:14:29 2020
 QLast Update : Wed Jul 29 07:14:29 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1060224	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	2661001	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.12	117	2440235	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.46	95	2020852	10.36	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.60%


Target Compounds	Ovalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL073020\
Data File : VL035270.D
Acq On : 30 Jul 2020 7:29
Operator : SY/AP
Sample : OC CAN 10601
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC CAN 10601

Quant Time: Aug 02 15:37:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL072920AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jul 29 07:14:29 2020
QLast Update : Wed Jul 29 07:14:29 2020
Response via : Initial Calibration





#1

Bromochloromethane

Concen: 10.000 ppbv

RT: 5.66 min Scan# 893

Delta R.T. -0.00 min

Lab File: VL035270.D

Acq: 30 Jul 2020 7:29

Instrument :

MSVOA_L

ClientSampleId :

QC CAN 10601


Tot Ion: 49 Resp: 1060224

Ion Ratio Lower Upper

49 100

128 50.7 24.0 72.0




130 65.5 30.7 92.1



Abundance Ion 49.10 (48.80 to 49.80): VL0

Ion 128.00 (127.70 to 128.70): VL0

Ion 130.00 (129.70 to 130.70): VL0



#33

1,4-Difluorobenzene

Concen: 10.000 ppbv

RT: 7.18 min Scan# 1366

Delta R.T. -0.01 min

Lab File: VL035270.D

Acq: 30 Jul 2020 7:29


Tgt Ion: 114 Resp: 2661001

Ion Ratio Lower Upper

114 100

63 27.5 22.2 33.2



88 19.2 15.4 23.0




Abundance Ion 114.10 (113.80 to 114.80): VL0

Ion 63.10 (62.80 to 63.80): VL0

Ion 88.10 (87.80 to 88.80): VL0









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 12.12 min Scan# 2902
Delta R.T. -0.01 min
Lab File: VL035270.D
Acq: 30 Jul 2020 7:29

Instrument :
MSVOA_L
ClientSampleId :
QC CAN 10601

Tot Ion:117 Resp: 2440235
Ion Ratio Lower Upper
117 100
82 68.6 52.4 78.6
119 33.4 24.6 36.8



#68
1-Bromo-4-Fluorobenzene
Concen: 10.364 ppbv
RT: 14.46 min Scan# 3629
Delta R.T. -0.00 min
Lab File: VL035270.D
Acq: 30 Jul 2020 7:29

Tgt Ion: 95 Resp: 2020852
Ion Ratio Lower Upper
95 100
174 65.2 52.7 79.1
175 4.7 3.9 5.9

