


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL073020\
 Data File : VL035263.D
 Acq On : 30 Jul 2020 2:09
 Operator : SY/AP
 Sample : BFB
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL072920AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Wed Jul 29 07:14:29 2020
 Last Update : Wed Jul 29 07:14:29 2020



AutoFind: Scans 3628, 3629, 3630; Background Corrected with Scan 3600

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	24.2	244245	PASS
75	95	30	66	55.5	559786	PASS
95	95	100	100	100.0	1009109	PASS
96	95	5	9	7.1	71504	PASS
173	174	0.00	2	0.7	4282	PASS
174	95	50	120	63.3	639253	PASS
175	174	4	9	7.5	47909	PASS
176	174	93	101	99.9	638656	PASS
177	176	5	9	7.0	44444	PASS