

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL073021\
 Data File : VL037293.D
 Acq On : 30 Jul 2021 20:31
 Operator : SY/AP
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

MMDadoda
 8/2/2021 8:57:47 AM

Quant Time: Jul 30 23:15:41 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL073021AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jul 30 2021
 QLast Update : Fri Jul 30 22:49:48 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1156551	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	3675983	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	3274148	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	2350227	9.57	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	95.70%

Target Compounds					Ovalue
5) Vinyl Chloride	3.25	62	3198m	0.039	ppbv
32) 1,1,1-Trichloroethane	6.51	97	7653m	0.032	ppbv
35) Carbon Tetrachloride	7.01	117	7005m	0.029	ppbv
38) Trichloroethene	7.83	130	4895m	0.034	ppbv
52) Tetrachloroethene	11.22	164	4369m	0.031	ppbv
63) 1,1,2,2-Tetrachloroethane	13.69	83	8300m	0.034	ppbv

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL073021\
Data File : VL037293.D
Acq On : 30 Jul 2021 20:31
Operator : SY/AP
Sample : VSTDICC0.03
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Manual Integrations
APPROVED

MMDadoda
8/2/2021 8:57:47 AM

Quant Time: Jul 30 23:15:41 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL073021AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jul
QLast Update : Fri Jul 30 22:49:48 2021
Response via : Initial Calibration

