

Data Path : W:\HPCHEM1\MSVOA_L\DATA\VL073115\
 Data File : VL025813.D
 Acq On : 31 Jul 2015 15:01
 Operator : SY/MD
 Sample : G3144-01DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 EFFLUENTDL

Manual Integrations
 APPROVED

sam
 8/4/2015 12:05:16 PM

Quant Time: Aug 03 05:12:05 2015
 Quant Method : W:\HPCHEM1\MSVOA_L\METHODS\VL070715AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jul 08 02:22:55 2015
 QLast Update : Wed Jul 08 02:22:55 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	6.67	49	474128	10.00	ppbv	0.01
33) 1,4-Difluorobenzene	8.34	114	1235906	10.00	ppbv	0.01
55) Chlorobenzene-d5	13.76	117	1574654	10.00	ppbv	0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	16.30	95	1431305	10.78	ppbv	0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	107.80%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
15) Acetone	4.65	43	30300	0.36	ppbv	# 72
20) Methylene Chloride	5.17	84	29419	0.86	ppbv	93
52) Tetrachloroethene	12.81	164	8199m	0.14	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA_L\DATA\VL073115\
Data File : VL025813.D
Acq On : 31 Jul 2015 15:01
Operator : SY/MD
Sample : G3144-01DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
EFFLUENTDL

Manual Integrations
APPROVED

sam
8/4/2015 12:05:16 PM

Quant Time: Aug 03 05:12:05 2015
Quant Method : W:\HPCHEM1\MSVOA_L\METHODS\VL070715AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jul
QLast Update : Wed Jul 08 02:22:55 2015
Response via : Initial Calibration

