

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL073120\
Data File : VL035297.D
Acq On : 31 Jul 2020 8:12
Operator : SY/AP
Sample : CAN 10282
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10282

Quant Time: Aug 01 03:52:33 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL072920AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jul 29 07:14:29 2020
QLast Update : Wed Jul 29 07:14:29 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1206996	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	2877892	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.12	117	2580012	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.46	95	2109882	10.23	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.30%


Target Compounds	Ovalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL073120\
Data File : VL035297.D
Acq On : 31 Jul 2020 8:12
Operator : SY/AP
Sample : CAN 10282
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10282

Quant Time: Aug 01 03:52:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL072920AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jul 29 07:14:29 2020
QLast Update : Wed Jul 29 07:14:29 2020
Response via : Initial Calibration





#1

Bromochloromethane

Concen: 10.000 ppbv

RT: 5.66 min Scan# 894

Delta R.T. 0.00 min

Lab File: VL035297.D

Acq: 31 Jul 2020 8:12

Instrument :

MSVOA_L

ClientSampleId :

CAN 10282


Tot Ion: 49 Resp: 1206996

Ion Ratio Lower Upper

49 100

128 48.7 24.0 72.0




130 63.0 30.7 92.1



Abundance Ion 49.10 (48.80 to 49.80): VL0

Ion 128.00 (127.70 to 128.70): VL0

Ion 130.00 (129.70 to 130.70): VL0



#33

1,4-Difluorobenzene

Concen: 10.000 ppbv

RT: 7.19 min Scan# 1368

Delta R.T. 0.00 min

Lab File: VL035297.D

Acq: 31 Jul 2020 8:12


Tgt Ion: 114 Resp: 2877892

Ion Ratio Lower Upper

114 100

63 27.1 22.2 33.2



88 18.9 15.4 23.0




Abundance Ion 114.10 (113.80 to 114.80): VL0

Ion 63.10 (62.80 to 63.80): VL0

Ion 88.10 (87.80 to 88.80): VL0









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 12.12 min Scan# 2903
Delta R.T. -0.00 min
Lab File: VL035297.D
Acq: 31 Jul 2020 8:12

Instrument :
MSVOA_L
ClientSampleId :
CAN 10282

Tot Ion:117 Resp: 2580012
Ion Ratio Lower Upper
117 100
82 68.5 52.4 78.6
119 33.5 24.6 36.8



#68
1-Bromo-4-Fluorobenzene
Concen: 10.234 ppbv
RT: 14.46 min Scan# 3631
Delta R.T. 0.00 min
Lab File: VL035297.D
Acq: 31 Jul 2020 8:12

Tgt Ion: 95 Resp: 2109882
Ion Ratio Lower Upper
95 100
174 64.8 52.7 79.1
175 4.7 3.9 5.9

