

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL073123\
 Data File : VL040570.D
 Acq On : 31 Jul 2023 13:11
 Operator : SY/MD
 Sample : VL0731ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0731ABL01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 08/01/2023
 Supervised By :Mahesh Dadoda 08/01/2023

Quant Time: Aug 01 01:38:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL071023AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 01 01:38:13 2023
 QLast Update : Mon Jul 10 16:32:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.185	49	1092705	10.000	ppbv	-0.02
33) 1,4-Difluorobenzene	6.645	114	2917487	10.000	ppbv	-0.02
55) Chlorobenzene-d5	11.462	117	2455127	10.000	ppbv	-0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.777	95	1897037	9.763	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.600%
Target Compounds						
						Qvalue
15) Acetone	3.478	43	24802m	0.197	ppbv	
20) Methylene Chloride	3.922	84	25060m	0.468	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL073123\
Data File : VL040570.D
Acq On : 31 Jul 2023 13:11
Operator : SY/MD
Sample : VL0731ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0731ABL01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 08/01/2023
Supervised By :Mahesh Dadoda 08/01/2023

Quant Time: Aug 01 01:38:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL071023AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Mon Jul 10 16:32:00 2023
Response via : Initial Calibration

