

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL080124\
 Data File : VL041472.D
 Acq On : 1 Aug 2024 10:38
 Operator : SY/MD
 Sample : QC073124 CAN 10321
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC073124 CAN 10321

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 08/02/2024
 Supervised By : Mahesh Dadoda 08/03/2024

Quant Time: Aug 02 04:11:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Aug 02 03:13:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.192	49	893085	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.652	114	2565626	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.465	117	2308692	10.000	ppbv	# 0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.774	95	1812097	10.199	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.000%
Target Compounds						
					Qvalue	
15) Acetone	3.497	43	11802	0.116	ppbv #	85
20) Methylene Chloride	3.935	84	7000m	0.185	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL080124\
Data File : VL041472.D
Acq On : 1 Aug 2024 10:38
Operator : SY/MD
Sample : QC073124 CAN 10321
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC073124 CAN 10321

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 08/02/2024
Supervised By :Mahesh Dadoda 08/03/2024

Quant Time: Aug 02 04:11:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Fri Aug 02 03:13:29 2024
Response via : Initial Calibration

