

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL080124\
 Data File : VL041480.D
 Acq On : 1 Aug 2024 16:17
 Operator : SY/MD
 Sample : P3445-01DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 1ST-FLOORDL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 08/02/2024
 Supervised By :Mahesh Dadoda 08/03/2024

Quant Time: Aug 02 04:12:22 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 02 04:12:22 2024

QLast Update : Fri Aug 02 03:13:29 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.195	49	938419	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.655	114	2483828	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.468	117	2259058	10.000	ppbv	# 0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.777	95	1730652	9.955	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.500%
Target Compounds						
						Qvalue
3) Chlorodifluoromethane	2.681	51	154286	1.381	ppbv	95
9) Propene	2.710	41	57085m	1.112	ppbv	
13) Ethanol	3.240	45	955665	208.352	ppbv	97
15) Acetone	3.475	43	433146	4.062	ppbv	98
19) Isopropyl Alcohol	3.581	45	456727m	7.327	ppbv	
20) Methylene Chloride	3.938	84	378226	9.513	ppbv	96
26) Hexane	5.214	57	175979	2.003	ppbv	# 44
34) 2-Butanone	4.793	43	21617m	0.180	ppbv	
53) Toluene	9.227	91	57173m	0.283	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL080124\
Data File : VL041480.D
Acq On : 1 Aug 2024 16:17
Operator : SY/MD
Sample : P3445-01DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
1ST-FLOORDL

Quant Time: Aug 02 04:12:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Aug 02 03:13:29 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 08/02/2024
Supervised By : Mahesh Dadoda 08/03/2024

