

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL080124\
 Data File : VL041482.D
 Acq On : 1 Aug 2024 17:42
 Operator : SY/MD
 Sample : P3445-02DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 AMBIENT-OUTSIDEDL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 08/02/2024
 Supervised By :Mahesh Dadoda 08/03/2024

Quant Time: Aug 02 04:13:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LFri Aug 2
 QLast Update : Fri Aug 02 03:13:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.195	49	887786	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.655	114	2353600	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.471	117	2146360	10.000	ppbv	# 0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.780	95	1611068	9.754	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.500%
Target Compounds						
					Qvalue	
9) Propene	2.703	41	11320m	0.233	ppbv	
13) Ethanol	3.253	45	37730m	8.695	ppbv	
15) Acetone	3.475	43	117338m	1.163	ppbv	
19) Isopropyl Alcohol	3.591	45	14428m	0.245	ppbv	
20) Methylene Chloride	3.938	84	460760	12.249	ppbv	98
26) Hexane	5.218	57	187312	2.253	ppbv	# 37

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL080124\
Data File : VL041482.D
Acq On : 1 Aug 2024 17:42
Operator : SY/MD
Sample : P3445-02DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
AMBIENT-OUTSIDEDL

Quant Time: Aug 02 04:13:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Aug 02 03:13:29 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 08/02/2024
Supervised By : Mahesh Dadoda 08/03/2024

