

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL080519\
 Data File : VL034097.D
 Acq On : 6 Aug 2019 9:49
 Operator : SY/AP
 Sample : K4098-01DL 2X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 E049-1100DL

Quant Time: Aug 06 15:08:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL080519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Aug 05 17:09:48 2019
 QLast Update : Mon Aug 05 17:09:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.71	49	1236806	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.24	114	2386248	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.18	117	2546122	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.52	95	2370442	11.47	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	114.70%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
8) Dichlorotetrafluoroethane	3.19	85	1363313	7.764	ppbv	94
11) Trichlorofluoromethane	3.97	101	1878389	10.408	ppbv	99
18) 1,1-Dichloroethene	4.31	96	541117	8.985	ppbv	94
24) 1,1-Dichloroethane	5.05	63	1867945	10.713	ppbv	99
29) Chloroform	5.80	83	2404421	12.004	ppbv	97
37) 1,2-Dichloroethane	6.36	62	1251774	8.579	ppbv	100
42) Bromodichloromethane	7.86	83	1922131	9.395	ppbv	98
58) Ethyl Benzene	12.80	91	4949588	11.561	ppbv	95
59) m/p-Xylene	13.06	91	2008378	5.844	ppbv	97
60) o-Xylene	13.79	91	2642217	7.693	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL080519\
Data File : VL034097.D
Acq On : 6 Aug 2019 9:49
Operator : SY/AP
Sample : K4098-01DL 2X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
E049-1100DL

Quant Time: Aug 06 15:08:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL080519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Aug 05 17:09:48 2019
QLast Update : Mon Aug 05 17:09:48 2019
Response via : Initial Calibration

