

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL080519\
 Data File : VL034093.D
 Acq On : 5 Aug 2019 18:07
 Operator : SY/AP
 Sample : VSTDICV010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL080519

Quant Time: Aug 05 19:36:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL080519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Aug 05 17:09:48 2019
 QLast Update : Mon Aug 05 17:09:48 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	108	0.00
2 T	Dichlorodifluoromethane	1.231	1.187	3.6	122	0.00
3	Chlorodifluoromethane	1.156	0.946	18.2	98	0.00
4	Chloromethane	0.745	0.614	17.6	97	0.00
5 T	Vinyl Chloride	0.657	0.554	15.7	96	0.00
6 T	Bromomethane	0.339	0.291	14.2	98	0.00
7	Chloroethane	0.246	0.201	18.3	97	0.00
8 T	Dichlorotetrafluoroethane	1.420	1.147	19.2	99	0.00
9 T	Propene	0.682	0.559	18.0	98	0.00
10 T	Heptane	1.763	1.442	18.2	97	0.00
11 T	Trichlorofluoromethane	1.459	1.180	19.1	98	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.030	0.833	19.1	98	0.00
13	Ethanol	0.192	0.096	50.0#	87	0.00
14 T	Bromoethene	0.460	0.380	17.4	96	0.00
15 T	Acetone	1.429	1.015	29.0	95	0.00
16 T	1,3-Butadiene	0.695	0.577	17.0	97	0.00
17	tert-Butyl alcohol	1.333	1.076	19.3	92	0.00
18 T	1,1-Dichloroethene	0.487	0.390	19.9	95	0.00
19 T	Isopropyl Alcohol	1.113	0.819	26.4	92	0.00
20 T	Methylene Chloride	0.429	0.344	19.8	97	0.00
21 T	Allyl Chloride	0.989	0.803	18.8	95	0.00
22 T	trans-1,2-Dichloroethene	0.473	0.384	18.8	94	0.00
23 T	Vinyl Acetate	2.369	1.872	21.0	94	0.00
24 T	1,1-Dichloroethane	1.410	1.135	19.5	95	0.00
25 T	Ethyl Acetate	2.811	2.196	21.9	95	0.00
26 T	Hexane	1.191	0.950	20.2	97	0.00
27 T	Carbon Disulfide	1.256	1.054	16.1	95	0.00
28 T	Methyl tert-Butyl Ether	1.992	1.588	20.3	95	0.00
29 T	Chloroform	1.620	1.330	17.9	98	0.00
30 T	Cyclohexane	0.941	0.791	15.9	99	0.00
31 T	cis-1,2-Dichloroethene	1.177	0.958	18.6	96	0.00
32 T	1,1,1-Trichloroethane	1.603	1.286	19.8	96	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	105	0.00
34 T	2-Butanone	0.930	0.734	21.1	95	0.00
35 T	Carbon Tetrachloride	0.768	0.605	21.2	95	0.00
36 T	Benzene	1.137	0.909	20.1	97	0.00
37 T	1,2-Dichloroethane	0.611	0.500	18.2	95	0.00
38 T	Trichloroethene	0.440	0.313	28.9	96	0.00
39 T	1,2-Dichloropropane	0.459	0.376	18.1	97	0.00
40 T	1,4-Dioxane	0.173	0.130	24.9	92	0.00
41 T	Tetrahydrofuran	0.572	0.468	18.2	96	0.00
42 T	Bromodichloromethane	0.857	0.708	17.4	96	0.00
43	Methyl Methacrylate	0.466	0.383	17.8	96	0.00
44 T	2,2,4-Trimethylpentane	2.035	1.564	23.1	98	0.00
45 T	t-1,3-Dichloropropene	0.591	0.504	14.7	94	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL080519\
 Data File : VL034093.D
 Acq On : 5 Aug 2019 18:07
 Operator : SY/AP
 Sample : VSTDICV010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICSVVL080519

Quant Time: Aug 05 19:36:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL080519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Aug 05 17:09:48 2019
 QLast Update : Mon Aug 05 17:09:48 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	cis-1,3-Dichloropropene	0.682	0.570	16.4	94	0.00
47 T	1,1,2-Trichloroethane	0.438	0.361	17.6	96	0.00
48 T	Dibromochloromethane	0.703	0.574	18.3	93	0.00
49 T	Bromoform	0.629	0.518	17.6	92	0.00
50 T	4-Methyl-2-Pentanone	1.415	1.110	21.6	96	0.00
51 T	2-Hexanone	1.189	0.950	20.1	96	0.00
52 T	Tetrachloroethene	0.439	0.309	29.6	92	0.00
53 T	Toluene	1.312	1.039	20.8	97	0.00
54 T	1,2-Dibromoethane	0.664	0.541	18.5	96	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
56	1,1,1,2-Tetrachloroethane	0.457	0.378	17.3	93	0.00
57 T	Chlorobenzene	0.901	0.725	19.5	95	0.00
58 T	Ethyl Benzene	1.682	1.359	19.2	96	0.00
59 T	m/p-Xylene	1.350	1.096	18.8	98	0.00
60 T	o-Xylene	1.349	1.077	20.2	97	0.00
61 T	Styrene	1.071	0.900	16.0	96	0.00
62	Isopropylbenzene	1.811	1.429	21.1	96	0.00
63 T	1,1,2,2-Tetrachloroethane	0.974	0.766	21.4	98	0.00
64	n-propylbenzene	0.479	0.392	18.2	96	0.00
65	tert-Butylbenzene	1.613	1.255	22.2	96	0.00
66 T	Benzyl Chloride	0.893	0.773	13.4	88	0.00
67	sec-Butylbenzene	2.276	1.759	22.7	97	0.00
68 S	1-Bromo-4-Fluorobenzene	0.811	0.849	-4.7	105	0.00
69	p-Isopropyltoluene	1.878	1.462	22.2	96	0.00
70	n-Butylbenzene	2.025	1.619	20.0	101	0.00
71	2-Chlorotoluene	1.456	1.176	19.2	98	0.00
72 T	4-Ethyltoluene	1.588	1.297	18.3	98	0.00
73 T	1,3,5-Trimethylbenzene	1.527	1.231	19.4	97	0.00
74 T	1,2,4-Trimethylbenzene	1.589	1.235	22.3	98	0.00
75 T	1,3-Dichlorobenzene	0.890	0.715	19.7	97	0.00
76 T	1,4-Dichlorobenzene	0.909	0.738	18.8	98	0.00
77 T	1,2-Dichlorobenzene	0.877	0.719	18.0	98	0.00
78 T	Hexachloro-1,3-Butadiene	0.695	0.613	11.8	115	0.00
79 T	Naphthalene	1.650	1.253	24.1	95	0.00
80 T	Naphthalene,2-methyl-	0.517	0.561	-8.5	140	0.00
81 T	1,2,4-Trichlorobenzene	0.916	0.713	22.2	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0