

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL080621\
 Data File : VL037342.D
 Acq On : 6 Aug 2021 12:34
 Operator : SY/AP
 Sample : VSTDIC0.1
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.1

Manual Integrations
 APPROVED

MMDadoda
 8/10/2021 6:10:32 PM

Quant Time: Aug 06 16:18:55 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL080621AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 06 2021
 QLast Update : Fri Aug 06 15:47:06 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	916991	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	3085306	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.07	117	2791359	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.40	95	1900418	9.74	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.25	62	9122m	0.129	ppbv	
32) 1,1,1-Trichloroethane	6.49	97	18506m	0.103	ppbv	
35) Carbon Tetrachloride	7.01	117	17783m	0.094	ppbv	
38) Trichloroethene	7.82	130	12210m	0.100	ppbv	
52) Tetrachloroethene	11.21	164	11105m	0.092	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.67	83	23676m	0.102	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL080621\
Data File : VL037342.D
Acq On : 6 Aug 2021 12:34
Operator : SY/AP
Sample : VSTDIC0.1
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.1

Manual Integrations
APPROVED

MMDadoda
8/10/2021 6:10:32 PM

Quant Time: Aug 06 16:18:55 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL080621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Aug 06 15:47:06 2021
Response via : Initial Calibration

