

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL081121\
 Data File : VL037379.D
 Acq On : 11 Aug 2021 5:00
 Operator : SY/AP
 Sample : VSTDICC0.1
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.1

Manual Integrations
 APPROVED

apatel
 8/12/2021 10:16:44 AM

Quant Time: Aug 11 06:08:45 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL081121AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug 11 06:08:45 2021
 QLast Update : Wed Aug 11 03:21:19 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1206783	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	3851328	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.08	117	3359725	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.41	95	2461997	9.72	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.20%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.26	62	8352m	0.097	ppbv	
32) 1,1,1-Trichloroethane	6.50	97	20962m	0.087	ppbv	
35) Carbon Tetrachloride	7.01	117	19624m	0.078	ppbv	
38) Trichloroethene	7.83	130	11864m	0.081	ppbv	
52) Tetrachloroethene	11.21	164	14051m	0.098	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.67	83	27879m	0.111	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL081121\
Data File : VL037379.D
Acq On : 11 Aug 2021 5:00
Operator : SY/AP
Sample : VSTDIC0.1
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.1

Manual Integrations
APPROVED

apatel
8/12/2021 10:16:44 AM

Quant Time: Aug 11 06:08:45 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL081121AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug
QLast Update : Wed Aug 11 03:21:19 2021
Response via : Initial Calibration

