

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL081219\
 Data File : VL034101.D
 Acq On : 12 Aug 2019 12:20
 Operator : SY/AP
 Sample : VL0812ABL01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0812ABL01

Quant Time: Aug 13 04:51:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL080519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Aug 05 17:09:48 2019
 QLast Update : Mon Aug 05 17:09:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.71	49	1269734	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.24	114	2400560	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.17	117	2465377	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.51	95	2141527	10.71	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	107.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.88	43	32067	0.177	ppbv	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL081219\
Data File : VL034101.D
Acq On : 12 Aug 2019 12:20
Operator : SY/AP
Sample : VL0812ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0812ABL01

Quant Time: Aug 13 04:51:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL080519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Aug 05 17:09:48 2019
QLast Update : Mon Aug 05 17:09:48 2019
Response via : Initial Calibration

