


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL081219\
Data File : VL034099.D
Acq On : 12 Aug 2019 9:53
Operator : SY/AP
Sample : BFB
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL080519AIR.M
Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Mon Aug 05 17:09:48 2019
Last Update : Mon Aug 05 17:09:48 2019



AutoFind: Scans 3598, 3599, 3600; Background Corrected with Scan 3576

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	25.4	233600	PASS
75	95	30	66	57.4	526613	PASS
95	95	100	100	100.0	917994	PASS
96	95	5	9	7.0	64410	PASS
173	174	0.00	2	0.8	4601	PASS
174	95	50	120	62.5	573525	PASS
175	174	4	9	7.8	44946	PASS
176	174	93	101	97.9	561386	PASS
177	176	5	9	6.5	36496	PASS