

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081224\
 Data File : VL041535.D
 Acq On : 12 Aug 2024 14:07
 Operator : SY/MD
 Sample : P3496-06DL2 200X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV-3DL2

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 08/13/2024
 Supervised By :Mahesh Dadoda 08/13/2024

Quant Time: Aug 13 02:13:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LFri Aug 2
 QLast Update : Tue Aug 13 02:09:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.198	49	960135	10.000	ppbv	0.01
33) 1,4-Difluorobenzene	6.655	114	2513402	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.468	117	2347282	10.000	ppbv	# 0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.777	95	1759698	9.742	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.400%
Target Compounds						
					Qvalue	
9) Propene	2.710	41	54298	1.034	ppbv	100
13) Ethanol	3.263	45	35585	7.583	ppbv	81
15) Acetone	3.475	43	1221887	11.201	ppbv	96
20) Methylene Chloride	3.941	84	17922	0.441	ppbv	98
26) Hexane	5.218	57	10071m	0.112	ppbv	
34) 2-Butanone	4.790	43	1679974	13.818	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081224\
Data File : VL041535.D
Acq On : 12 Aug 2024 14:07
Operator : SY/MD
Sample : P3496-06DL2 200X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV-3DL2

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 08/13/2024
Supervised By :Mahesh Dadoda 08/13/2024

Quant Time: Aug 13 02:13:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Tue Aug 13 02:09:20 2024
Response via : Initial Calibration

