

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081224\
 Data File : VL041527.D
 Acq On : 12 Aug 2024 8:46
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Aug 13 02:09:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Aug 13 02:09:20 2024
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	77	0.00
2 T	Dichlorodifluoromethane	1.720	1.051	38.9#	57	0.00
3	Chlorodifluoromethane	1.190	1.200	-0.8	87	0.00
4	Chloromethane	0.633	0.653	-3.2	89	0.00
5 T	Vinyl Chloride	0.624	0.655	-5.0	89	0.00
6 T	Bromomethane	0.321	0.342	-6.5	90	0.00
7	Chloroethane	0.215	0.222	-3.3	85	0.00
8 T	Dichlorotetrafluoroethane	1.524	1.495	1.9	84	0.00
9 T	Propene	0.547	0.547	0.0	86	0.00
10 T	Heptane	1.229	1.374	-11.8	89	0.00
11 T	Trichlorofluoromethane	1.528	1.540	-0.8	86	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.100	1.123	-2.1	87	0.00
13	Ethanol	0.049	0.042#	14.3	92	0.00
14 T	Bromoethene	0.429	0.448	-4.4	84	0.00
15 T	Acetone	1.136	1.170	-3.0	91	0.00
16 T	1,3-Butadiene	0.607	0.644	-6.1	90	0.00
17	tert-Butyl alcohol	1.036	1.233	-19.0	91	0.00
18 T	1,1-Dichloroethene	0.471	0.495	-5.1	84	0.00
19 T	Isopropyl Alcohol	0.664	0.728	-9.6	87	0.00
20 T	Methylene Chloride	0.424	0.413	2.6	86	0.00
21 T	Allyl Chloride	0.745	0.814	-9.3	87	0.00
22 T	trans-1,2-Dichloroethene	0.471	0.527	-11.9	91	0.00
23 T	Vinyl Acetate	0.868	1.041	-19.9	91	0.00
24 T	1,1-Dichloroethane	1.022	1.107	-8.3	91	0.00
25 T	Ethyl Acetate	2.257	2.460	-9.0	90	0.00
26 T	Hexane	0.936	1.018	-8.8	89	0.00
27 T	Carbon Disulfide	1.095	1.211	-10.6	84	0.00
28 T	Methyl tert-Butyl Ether	0.591	0.654	-10.7	91	0.00
29 T	Chloroform	1.527	1.598	-4.6	87	0.00
30 T	Cyclohexane	0.763	0.870	-14.0	88	0.00
31 T	cis-1,2-Dichloroethene	0.925	0.997	-7.8	87	0.00
32 T	1,1,1-Trichloroethane	1.487	1.574	-5.9	87	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	78	0.00
34 T	2-Butanone	0.484	0.530	-9.5	88	0.00
35 T	Carbon Tetrachloride	0.552	0.600	-8.7	85	0.00
36 T	Benzene	0.731	0.789	-7.9	87	0.00
37 T	1,2-Dichloroethane	0.409	0.442	-8.1	87	0.00
38 T	Trichloroethene	0.335	0.365	-9.0	88	0.00
39 T	1,2-Dichloropropane	0.280	0.304	-8.6	89	0.00
40 T	1,4-Dioxane	0.112	0.121	-8.0	82	0.00
41 T	Tetrahydrofuran	0.277	0.323	-16.6	90	0.00
42 T	Bromodichloromethane	0.591	0.642	-8.6	86	0.00
43	Methyl Methacrylate	0.262	0.311	-18.7	85	0.00
44 T	2,2,4-Trimethylpentane	1.220	1.331	-9.1	89	0.00
45 T	t-1,3-Dichloropropene	0.241	0.321	-33.2#	87	0.00
46 T	cis-1,3-Dichloropropene	0.348	0.430	-23.6	87	0.00
47 T	1,1,2-Trichloroethane	0.308	0.326	-5.8	86	0.00
48 T	Dibromochloromethane	0.497	0.552	-11.1	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081224\
 Data File : VL041527.D
 Acq On : 12 Aug 2024 8:46
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Aug 13 02:09:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Aug 13 02:09:20 2024
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.420	0.478	-13.8	83	0.00
50 T	4-Methyl-2-Pentanone	0.763	0.877	-14.9	88	0.00
51 T	2-Hexanone	0.577	0.716	-24.1	87	0.00
52 T	Tetrachloroethene	0.282	0.298	-5.7	85	0.00
53 T	Toluene	0.814	0.923	-13.4	86	0.00
54 T	1,2-Dibromoethane	0.410	0.478	-16.6	86	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	79	0.00
56	1,1,1,2-Tetrachloroethane	0.388	0.408	-5.2	87	0.00
57 T	Chlorobenzene	0.712	0.728	-2.2	86	0.00
58 T	Ethyl Benzene	1.153	1.253	-8.7	86	0.00
59 T	m/p-Xylene	0.995	1.068	-7.3	86	0.00
60 T	o-Xylene	0.941	1.022	-8.6	87	0.00
61 T	Styrene	0.425	0.499	-17.4	85	0.00
62	Isopropylbenzene	1.431	1.483	-3.6	86	0.00
63 T	1,1,2,2-Tetrachloroethane	0.594	0.623	-4.9	87	0.00
64	n-propylbenzene	0.374	0.402	-7.5	86	0.00
65	tert-Butylbenzene	1.292	1.341	-3.8	86	0.00
66 T	Benzyl Chloride	0.083	0.108	-30.1#	93	0.00
67	sec-Butylbenzene	1.781	1.875	-5.3	86	0.00
68 S	1-Bromo-4-Fluorobenzene	0.770	0.725	5.8	73	0.00
69	p-Isopropyltoluene	1.478	1.578	-6.8	86	0.00
70	n-Butylbenzene	1.515	1.632	-7.7	87	0.00
71	2-Chlorotoluene	1.055	1.126	-6.7	86	0.00
72 T	4-Ethyltoluene	1.108	1.230	-11.0	85	0.00
73 T	1,3,5-Trimethylbenzene	0.955	1.052	-10.2	85	0.00
74 T	1,2,4-Trimethylbenzene	1.110	1.178	-6.1	87	0.00
75 T	1,3-Dichlorobenzene	0.698	0.731	-4.7	86	0.00
76 T	1,4-Dichlorobenzene	0.676	0.712	-5.3	86	0.00
77 T	1,2-Dichlorobenzene	0.692	0.711	-2.7	86	0.00
78 T	Hexachloro-1,3-Butadiene	0.590	0.539	8.6	86	0.00
79 T	Naphthalene	0.980	1.197	-22.1	84	0.00
80 T	Naphthalene,2-methyl-	0.821	1.016	-23.8	81	0.00
81 T	1,2,4-Trichlorobenzene	0.583	0.631	-8.2	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0