

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081518\
 Data File : VL032353.D
 Acq On : 15 Aug 2018 13:15
 Operator : SY/AP
 Sample : VSTDICC0.03
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

MMDadoda
 8/17/2018 2:06:32 PM

Quant Time: Aug 15 14:33:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081518AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug 15 2018
 QLast Update : Wed Aug 15 11:44:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.82	49	2243497	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.35	114	4781457	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.30	117	4537005	10.00	ppbv	-0.02

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
68) 1-Bromo-4-Fluorobenzene	14.64	95	3499944	9.94	ppbv	-0.03	
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.40%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Vinyl Chloride	3.34	62	3943	0.032	ppbv	# 70
32) 1,1,1-Trichloroethane	6.68	97	9463m	0.034	ppbv	
35) Carbon Tetrachloride	7.19	117	7867m	0.029	ppbv	
38) Trichloroethene	8.02	130	5489m	0.034	ppbv	
52) Tetrachloroethene	11.45	164	3402m	0.024	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.91	83	12418m	0.032	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081518\
Data File : VL032353.D
Acq On : 15 Aug 2018 13:15
Operator : SY/AP
Sample : VSTDICC0.03
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Manual Integrations
APPROVED

MMDadoda
8/17/2018 2:06:32 PM

Quant Time: Aug 15 14:33:16 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081518AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug
QLast Update : Wed Aug 15 11:44:08 2018
Response via : Initial Calibration

