

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL081518\
 Data File : VL032360.D
 Acq On : 15 Aug 2018 20:31
 Operator : SY/AP
 Sample : J4266-01DL 4X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 1000DL

Quant Time: Aug 16 22:13:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL081518AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Wed Aug 15 14:58:02 2018
 QLast Update : Wed Aug 15 14:58:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.82	49	2389740	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.35	114	4632543	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.30	117	4436798	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.65	95	3549575	10.49	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.33	62	549348	4.37	ppbv	99
29) Chloroform	5.90	83	1648914	5.19	ppbv	97
36) Benzene	7.06	78	1310859	3.34	ppbv	98
37) 1,2-Dichloroethane	6.47	62	702288	2.95	ppbv	100
39) 1,2-Dichloropropane	7.80	63	849539	5.18	ppbv	98
52) Tetrachloroethene	11.44	164	459252	3.96	ppbv	93
54) 1,2-Dibromoethane	10.83	107	1395526	6.02	ppbv	98
57) Chlorobenzene	12.37	112	1078859	3.23	ppbv	99
60) o-Xylene	13.92	91	1087431	2.44	ppbv	98
77) 1,2-Dichlorobenzene	18.06	146	1269247	4.34	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL081518\
Data File : VL032360.D
Acq On : 15 Aug 2018 20:31
Operator : SY/AP
Sample : J4266-01DL 4X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
1000DL

Quant Time: Aug 16 22:13:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL081518AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug 15 14:58:02 2018
QLast Update : Wed Aug 15 14:58:02 2018
Response via : Initial Calibration

