

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL081518\
 Data File : VL032361.D
 Acq On : 15 Aug 2018 21:12
 Operator : SY/AP
 Sample : J4266-01DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 1000DL

Quant Time: Aug 16 06:58:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL081518AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Wed Aug 15 14:58:02 2018
 QLast Update : Wed Aug 15 14:58:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.82	49	2420310	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.35	114	4685332	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.31	117	4483849	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.64	95	3505273	10.25	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.33	62	207770	1.63	ppbv	96
29) Chloroform	5.90	83	665662	2.07	ppbv	97
36) Benzene	7.06	78	461127	1.16	ppbv	99
37) 1,2-Dichloroethane	6.47	62	265624	1.10	ppbv	100
39) 1,2-Dichloropropane	7.80	63	317588	1.91	ppbv	97
52) Tetrachloroethene	11.44	164	162177	1.38	ppbv	92
54) 1,2-Dibromoethane	10.82	107	494285	2.11	ppbv	97
57) Chlorobenzene	12.37	112	415838	1.23	ppbv #	92
60) o-Xylene	13.92	91	304464	0.68	ppbv	98
77) 1,2-Dichlorobenzene	18.06	146	432326	1.46	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL081518\
Data File : VL032361.D
Acq On : 15 Aug 2018 21:12
Operator : SY/AP
Sample : J4266-01DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
1000DL

Quant Time: Aug 16 06:58:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL081518AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug 15 14:58:02 2018
QLast Update : Wed Aug 15 14:58:02 2018
Response via : Initial Calibration

