

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL081720\
 Data File : VL035441.D
 Acq On : 18 Aug 2020 12:02
 Operator : SY/AP
 Sample : L3510-01DL 4X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 E053-1100DL

Quant Time: Aug 19 17:03:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL081720AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Mon Aug 17 19:28:30 2020
 QLast Update : Mon Aug 17 19:28:30 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1036787	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	4492449	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.11	117	4491988	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.44	95	3242701	10.31	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
6) Bromomethane	3.47	94	745342	10.530	ppbv	99
18) 1,1-Dichloroethene	4.29	96	776486	7.827	ppbv #	59
29) Chloroform	5.76	83	1596782	6.065	ppbv	99
38) Trichloroethene	7.85	130	1763170	9.613	ppbv #	82
42) Bromodichloromethane	7.81	83	2857081	11.331	ppbv #	98
47) 1,1,2-Trichloroethane	9.49	97	1198538	7.384	ppbv	92
48) Dibromochloromethane	10.32	129	2165242	7.952	ppbv	99
58) Ethyl Benzene	12.74	91	3236945	5.969	ppbv #	89
76) 1,4-Dichlorobenzene	17.17	146	4344699	9.657	ppbv	91
77) 1,2-Dichlorobenzene	17.85	146	4153559	9.289	ppbv	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL081720\
Data File : VL035441.D
Acq On : 18 Aug 2020 12:02
Operator : SY/AP
Sample : L3510-01DL 4X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
E053-1100DL

Quant Time: Aug 19 17:03:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL081720AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Aug 17 19:28:30 2020
QLast Update : Mon Aug 17 19:28:30 2020
Response via : Initial Calibration

