


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL081720\
 Data File : VL035419.D
 Acq On : 17 Aug 2020 13:31
 Operator : SY/AP
 Sample : BFB
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081720AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Mon Aug 17 19:28:30 2020
 Last Update : Mon Aug 17 19:28:30 2020



AutoFind: Scans 3625, 3626, 3627; Background Corrected with Scan 3600

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	14.8	227178	PASS
75	95	30	66	46.6	714261	PASS
95	95	100	100	100.0	1532074	PASS
96	95	5	9	7.1	109405	PASS
173	174	0.00	2	0.5	7502	PASS
174	95	50	120	95.4	1461418	PASS
175	174	4	9	7.6	111365	PASS
176	174	93	101	98.0	1432576	PASS
177	176	5	9	7.2	102898	PASS