

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081722\
 Data File : VL039541.D
 Acq On : 17 Aug 2022 6:51
 Operator : SY/AP
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 08/18/2022
 Supervised By : Mahesh Dadoda 08/19/2022

Quant Time: Aug 17 07:48:09 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081722AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug :
 QLast Update : Wed Aug 17 04:35:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.204	49	1064893	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.660	114	2756811	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.471	117	2519919	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.776	95	1901927	10.165	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.700%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	2.944	62	1664m	0.023	ppbv	
32) 1,1,1-Trichloroethane	6.014	97	3767m	0.021	ppbv	
35) Carbon Tetrachloride	6.506	117	4815m	0.029	ppbv	
38) Trichloroethene	7.300	130	3308m	0.029	ppbv	
52) Tetrachloroethene	10.625	164	2236m	0.026	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.043	83	3714m	0.027	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081722\
Data File : VL039541.D
Acq On : 17 Aug 2022 6:51
Operator : SY/AP
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Quant Time: Aug 17 07:48:09 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081722AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Aug
QLast Update : Wed Aug 17 04:35:27 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 08/18/2022
Supervised By :Mahesh Dadoda 08/19/2022

