

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081722\
 Data File : VL039546.D
 Acq On : 17 Aug 2022 12:45
 Operator : SY/AP
 Sample : N3986-01DL 4X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 E061-1100DL

Quant Time: Aug 18 05:36:57 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081722AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Aug 17 12:06:24 2022
 QLast Update : Wed Aug 17 12:06:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.204	49	1047517	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.661	114	2753351	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.471	117	2486502	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.776	95	1962503	10.484	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.800%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.751	85	853351	6.061	ppbv	98
5) Vinyl Chloride	2.944	62	317845	4.910	ppbv	98
12) 1,1,2-Trichlorotrifluo...	4.085	101	925746	8.103	ppbv	99
29) Chloroform	5.285	83	1165300	7.195	ppbv	96
36) Benzene	6.378	78	813525	3.877	ppbv	97
42) Bromodichloromethane	7.252	83	977340	6.134	ppbv	99
53) Toluene	9.223	91	1644332	6.994	ppbv	99
58) Ethyl Benzene	12.091	91	1637250	5.686	ppbv	100
73) 1,3,5-Trimethylbenzene	15.355	105	2570264	9.651	ppbv	97
76) 1,4-Dichlorobenzene	16.474	146	1824501	10.099	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL081722\
Data File : VL039546.D
Acq On : 17 Aug 2022 12:45
Operator : SY/AP
Sample : N3986-01DL 4X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
E061-1100DL

Quant Time: Aug 18 05:36:57 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL081722AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Aug 17 12:06:24 2022
QLast Update : Wed Aug 17 12:06:24 2022
Response via : Initial Calibration

